



क्षेत्रीय निदेशालय ( मध्य ), भोपाल  
Regional Directorate (Central), Bhopal  
केन्द्रीय प्रदूषण नियंत्रण बोर्ड  
Central Pollution Control Board

( पर्यावरण, वन और जलवायु परिवर्तन मंत्रालय, भारत सरकार )  
( Ministry of Environment, Forests & Climate Change,  
Government of India )

क्षे.नि.भो/लैब(सीओडी डाईजेस्टर)/NIQ/2019-20/ 2096  
प्रति,

दिनांक 17 फरवरी, 2020

सूची संलग्न

विषय: प्रयोगशाला उपयोग हेतु सीओडी डाईजेस्टर प्रदाय करने बाबत मुहरबंद कोटेशन आमंत्रण सूचना।

केन्द्रीय प्रदूषण नियंत्रण बोर्ड, क्षेत्रीय निदेशालय (मध्य), भोपाल द्वारा प्रयोगशाला उपयोग हेतु सीओडी डाईजेस्टर प्रदाय करने बाबत मुहरबंद कोटेशन आमंत्रित किए जाते हैं, जिसकी तकनीकी विशेषताएं अनुलग्नक-1 पर संलग्न हैं।

उपरोक्त मुहरबंद कोटेशन निम्नलिखित दस्तावेजों के साथ प्रस्तुत करने पर ही स्वीकार किए जायेंगे :-

1. निर्माता द्वारा जारी वैध अधिकृत विक्रेता/डीलर का प्रमाण पत्र।
2. उपकरण हेतु कम से कम तीन वर्ष की कम्प्रेहेंसिव वारंटी/गारंटी प्रमाण पत्र।
3. GST संख्या का वैध प्रमाण पत्र।

सभी उल्लेखित नियम एवं शर्तों के साथ मुहरबंद लिफाफे, लिफाफे के उपर स्वच्छ एवं स्पष्ट शब्दों में "Scaled Quotation for Supply of COD Digester" केन्द्रीय प्रदूषण नियंत्रण बोर्ड, क्षेत्रीय निदेशालय (मध्य), 'परिवेश भवन', ई-5, पर्यावरण परिसर, अरेरा कालोनी, भोपाल-462016 के नाम पर दिनांक 06 मार्च, 2020 तक या उससे पूर्व कार्यालय में जमा किए जा सकते हैं। कोई भी कारण बताएं किसी भी कोटेशन को स्वीकार/निरस्त/अमान्य करने का सर्वाधिकार सक्षम अधिकारी को निहित है।

भवदीय

/  
(पी. जगन)  
क्षेत्रीय निदेशक

प्रतिलिपि:

1. प्रबंधन प्रतिनिधि एवं प्रयोगशाला प्रभारी, क्षेत्रीय निदेशालय, भोपाल - कृपया सूचनार्थ।
2. लेखा अधिकारी, क्षेत्रीय निदेशालय, भोपाल - कृपया सूचनार्थ।
3. प्रमुख सूचना प्रौद्योगिकी अनुभाग, के.प्र.नि.बो., दिल्ली - कृपया सूचनार्थ एवं वेबसाइट पर अपलोड करने हेतु।
4. सूचना पटल - कृपया सूचनार्थ।

  
क्षेत्रीय निदेशक

तृतीय तल, सहकार भवन, नॉर्थ टी.टी. नगर, भोपाल-462003 (म.प्र.)

फैक्स : 0755-2775587, ईपीएबीएक्स : 2775385, 2775386, Website : www.cpcb.nic.in, E-mail : cpcb.bhopal@gmail.com -

Head Office : Parivesh Bhawan, East Arjun Nagar, Delhi - 110 032

Tel.: (011) 43102030, Fax : 011-22304948, 22307078 223 707, Website : www.cpcb.nic.in



क्षेत्रीय निदेशालय ( मध्य ), भोपाल  
Regional Directorate (Central), Bhopal  
केन्द्रीय प्रदूषण नियंत्रण बोर्ड  
Central Pollution Control Board  
( पर्यावरण, वन और जलवायु परिवर्तन मंत्रालय, भारत सरकार )  
(Ministry of Environment, Forests & Climate Change,  
Government of India)

RD/BPL/Lab(COD Digester)/NIQ/2019-20/ 2098

Date: 17 February, 2020

To,

List Enclosed.

Sub: Notice for inviting sealed quotations (NIQ) for Supply of COD Digester for laboratory.

Central Pollution Control Board, Regional Directorate (Central), Bhopal invites Sealed quotations for Supply of COD Digester for laboratory as per Technical Speciation enclosed at Annexure-1.

The Sealed quotation shall only be accepted which are attached with following documents:

1. Valid authorized seller / dealer certificate from the Manufactures.
2. Minimum 3 Years Comprehensive Warranty / Guarantee for equipment.
3. Certificate for Valid GST number.

The Sealed quotations along with the term & conditions in sealed cover super scribed as "**Sealed Quotation for Supply of COD Digester**" may be submitted in the office address : Central Pollution Control Board, Regional Directorate (Central), 'Parivesh Bhawan', E-5, Paryavaran Parisar, Arera Colony, Bhopal- 462016 on or before 06 March, 2020. The right of cancellation/acceptation/rejection of any quotation is reserved with the competent authority without assigning any reason thereof.

1  
(P.Jagan)  
Regional Director

- Copy to: 1. M.R & Lab Incharge., CPCB, Bhopal --- for information please.  
2. Account Assistant, CPCB, Bhopal --- for information please.  
3. IT Division, CPCB, Delhi - for information and uploading at office website please.  
4. Display Board - for information please.


  
Regional Director


**TECHNICAL SPECIFICATION FOR COD DIGESTION APPARATUS WITH ALL ACCESSORIES**

**Requirement :**


- (A) Heating Block : Compact and Solid Metal block with 15 holes
- (B) Glassware : Reaction Vessels (dia 40 mm Approx.) : 15 Nos.  
& compatible Air Condensers – 15 Nos.
- (C) Temperature :  $150^{\circ}\text{C} \pm 1\%$  (Uniform heating from all holes)
- (D) Capacity : Analysis of 15 samples at a time.
- (E) Vessel Capacity : 200 ml (without neck)
- (F) Display – Digital : Graphic LCD Module 128x64 dots with large Font
- (G) Key board : Soft touch key board
- (H) Timer : Programmable Timer with Alarm
- (I) Power Supply : 230 Volt, 50 Hz, Single phase
- (J) Handling Kit / Water Tank : For Glass ware Handling
- (K) Electric Power Cable, Tool Kit, etc.
- (L) Comprehensive Warrantee : 03 years (Three Years)


Prepared By:

  
(Rajeev Sharma)  
Sr. Tech.

  
(Rashmi Thakur)  
SSA

Verified By:

  
(Dr. R.P. Mishra)  
Sci. 'D' & Lab Incharge

  
(P. Jagan)  
Regional Director