



केन्द्रीय प्रदूषण नियंत्रण बोर्ड
CENTRAL POLLUTION CONTROL BOARD
(पर्यावरण एवं वन मंत्रालय, भारत सरकार)
(MINISTRY OF ENVIRONMENT & FORESTS, GOVT. OF INDIA)

C-48011/01/2012-13/Mat.

Date: 20.06.2012


Sub:- Quotation for the supply of pH Tester and Conductivity Meters (Pen type low & high range)

Sir,

Sealed quotations are invited for the supply of **pH Tester and Conductivity Meters (Pen type low & high range)** as per specifications attached at **Annexure – I**

01. The quotation should reach this office on or before 16.07.2012.
02. The rate quoted should be valid till 31.12.2012 and as per our specification only.
03. The quotation should be sent by registered post/speed post. However, quotation may be put in the tender box kept in the first floor, material section, Parivesh Bhawan to avoid any delay.
04. It may be noted that mere quoting of lowest rates will not entitle any firm to get the order.
05. All the terms and conditions such as taxes/Vat, validity, guarantee period and transportation charges etc. should be clearly mentioned in the quotations. The quotation should be accompanied with technical literature.
06. The rate should be written both in word & figures and there should not be any variations in the same. In case of any variation in the price quoted, the price mentioned in words will be considered as final.
07. Payment Terms: 100% payment will be released on supply of the items in good condition. No advance payment will be made. Hence no preference will be given for such advance payment proposals.
08. The items should be supplied within four weeks from the date of issue of confirmed order. Failure to supply within the specified time, the order is liable to be cancelled.
09. You are requested to quote the price of Material as FOR Central Pollution Control Board and delivery will be made in the concerned sections.
10. The Chairman, Central Pollution Control Board reserves the right to cancel any or all tenders without assigning any reason thereof. Your quotations should be valid for one year from the date of its opening.
11. The All correspondence in this regard may be addressed to the Administrative Officer (Materials), Central Pollution Control Board, Delhi – 110 032.
12. Typographical and clerical errors are subject corrections.
13. In the event of any dispute, the decision of the Chairman, Central Board shall be final and binding upon both the parties.

Yours faithfully,


(M. Varghese)
Admn. Officer (M)

'परिवेश भवन' पूर्वी अर्जुन नगर, दिल्ली-110032

'Parivesh Bhawan', East Arjun Nagar, Delhi - 110032

दूरभाष/Tel. : 43102030, फ़ैक्स/Fax : 22305793, 22307078, 22307079, 22301932, 22304948

ई-मेल/e-mail : cpcb@nic.in वेबसाइट/Website : www.cpcb.nic.in

TECHNICAL SPECIFICATION

pH TESTER PEN TYPE – 3 NOS.

pH Tester, pen type

1. Range: 0-14 .0 pH
2. Resolution: 0.1 pH
3. Accuracy: ± 0.1 pH
4. Automatic Temp. Compensation 0 to 50⁰ C
5. Calibration Manual, Two Point

Other

6. Operating temp.: 0 to 50⁰ C
7. Power Supply: Battery Cell (tablet)
8. Calibration Solution:
 1. pH-4 100 ml (NIST Certified)
 2. pH-7 100 ml (NIST Certified)
 3. pH-9.2 100 ml (NIST Certified)

CONDUCTIVITY METER PEN TYPE (LOW RANGE) – 2 NOS.

Power Supply: Battery cell (tablet) operated

Range : 1 – 1999 umhos/cm

Accuracy : +/-2% FS

Calibration : manual, one point

Temp. Compensation : automatic 0 to 50⁰ C

FEATURES:

Auto Temps. Compensation

Resolution : 0.1 umhos/cm

Accuracy : ± 0.2 umhos/cm

Calibration Solution Pouch : 1413 us/cm calibration solution

CONDUCTIVITY METER PEN TYPE (HIGH RANGE) – 1 NOS.

Power Supply: Battery cell (tablet) operated

Range : 10 – 19990 umhos/cm

FEATURES:

Auto Temps. Compensation

Resolution : 10 umhos/cm

Accuracy : +/-2% FS

Calibration : manual, one point

Temp. Compensation : automatic 0 to 50⁰ C

Calibration Solution Pouch : 5.0 mS/cm calibration solution, 20 ml sachets

12.88 mS/cm Calibration Solution

