



केन्द्रीय प्रदूषण नियंत्रण बोर्ड

(पर्यावरण एवं वन मंत्रालय, भारत सरकार)

**CENTRAL POLLUTION CONTROL BOARD**

(MINISTRY OF ENVIRONMENT & FORESTS, GOVT. OF INDIA)

F.No.: Tech-42/AirLab/P-II(13)ZOB/13-14/

6<sup>th</sup> January 2014

To

As per the list

Sub: Inviting quotations for supplying of PM<sub>2.5</sub> sampler, Micro Analytical Balance, UV-Vis Spectrophotometer, Pyrometer and Personal VOC sampler

Sir,

Sealed quotations are hereby invited from the reputed manufacturers/Authorized dealers to supply the instruments/equipment namely PM<sub>2.5</sub> sampler, Micro Analytical Balance, UV-Vis Spectrophotometer, Pyrometer and Personal VOC sampler as per the specifications mentioned in Annexures and as per the terms and conditions detailed below:-

1. Technical bid containing the technical specification of the instruments/equipment shall be kept in a separate sealed envelope (1) and Price bid quoting the price of the instruments/equipment shall be kept separately in another sealed envelope (2) mentioning "Technical Bid" and "Price bid" on top left side of the envelope respectively.
2. Both Technical and Price bids are to be kept separately in sealed envelopes and both shall be kept in another sealed envelope(3) mentioning that "Quotation for instruments/equipment (Air)" on top left side of the envelope.
3. Separate quotation shall be submitted for each and every instrument/equipment individually depending on capacity to supply the instrument/equipment as per the terms and conditions.
4. Combined submission of the quotations for all the instrument/equipment shall be summarily rejected.

contd... /  
S. Swamy

दक्षिण आंचलिक कार्यालय : निसर्ग भवन, ए-ब्लॉक, प्रथम एवं द्वितीय तल, तिम्मय्या रोड, ७-डी मेन, शिवनगर, बंगलूरु - ५६० ०१०.

Zonal Office (South) : " Nisarga Bhawan ", A-Block, 1<sup>st</sup> & 2<sup>nd</sup> Floors, Thimmaiah Road, 7<sup>th</sup> D - Main, Shivanagar, Bangalore - 560 010.

दूरभाष / Telephone : 80-23233739. 23233827, 23233996, 23233600, 23232559, 23226002, 23222539, Fax : 080-23234059

ई-मेल / E-mail : cpcbszo@yahoo.com

प्रधान कार्यालय : परिवेश भवन, पूर्वी अर्जुन नगर, दिल्ली- ११० ०३२.

Head Office : Parivesh Bhawan, East Arjun Nagar, Delhi - 110 032.

दूरभाष / Telephone : 011-43102030, Fax : 22305793, 22307078, 22307079, 22301932, 22304948

ई-मेल / E-mail : cpcb@nic.in वेबसाइट / Website : www.cpcb.nic.in

5. Buy/Back : Quotation for buy/back shall also be quoted and preference will be given for the same. The instruments/equipment meant for buy/back shall be inspected during the Office hours at our premises if required.
6. The duly filled in quotation shall reach this office on or before 29<sup>th</sup> January 2014 at 16:00 hours by post. The quotation may also be dropped in the Tender Box kept at this office.
7. The quotation received without separate sealed envelopes for technical and Price bid as explained in conditions 1 and 2 shall not be entertained.
8. The instruments/equipment shall be delivered, installed and adequate demo & training to the laboratory staff shall be given at this office.
9. On site warranty for three years to be given.
10. The call of quotation does not bind the Central Board to place the order.
11. The quotations submitted in response to this notice may be rejected without assigning any reason.
12. The instruments/equipment shall be delivered along with the invoice in original at this office within 30 days from the date of issue of Supply Order
13. The validity of quotation shall normally be 90 days from the date of opening.
14. The payment shall be released after the delivery, demo and satisfactory performance of the instruments/equipment.
15. Applicable taxes, cess and any other charges if any shall be quoted separately.

Yours faithfully

Sd.../  
(S.Suresh)  
ZONAL OFFICER

Encl.: Annexure

Copy to

1. I/c. IT Divn., CPCB, HO. With a request to place in the official website
2. Notice Board, CPCB, SZO., Bangalore
3. Notice Board, KSPCB, Bangalore

S-Suresh  
6/1/2014

### Micro Analytical Balance

S. No.	Parameters	Specifications	Tick mark should be put as applicable (Either <input checked="" type="checkbox"/> or <input type="checkbox"/> )
1	Capacity, Max	0.00001mg to 5.1 or 6.1 g	
2	Readability	± 1 µg	
3	Repeatability	± 1 µg	
4	Linearity, micrograms	± 4 µg	
5	Linearity	0 - 500 µg ± 2 µg	
6	Response time	10 sec	
7	Operating Temperature	5 °C to 40 °C	
8	Operating Humidity	85% RH	
9	Zero tracking	Automatic, with Level Adjustment	
10	Calibration	Self Calibration	
11	Power Source	230 ± 10V AC with suitable AC adopter	
12	Standard Features	1. 230 ± 10V AC Adapter, 2. Level indicator, with adjustable legs, 3. RS-232, /USB Port 4. Operational manual	
13	Warranty	Minimum three years warranty from the date of installation	
14	Calibration Certificate	From Accredited Agency with one year validity	
15	Shield	Suitable Size of Draft Shield	
16	Built in Interface	USB/ RS - 232	

Note: Furnish two quotes    1) Regular Price  
    2) Buy back price

*S. Suman*

*[Handwritten Signature]*