

SECOND CALL

F.No.: Tech-42/AirLab/P-II(13)ZOB/13-14/

19th February, 2014

To

As per the list

Sub: Inviting quotations for supplying of Micro Analytical Balance

Sir,

Sealed quotations are hereby invited from the reputed manufacturers/Authorized dealers to supply Micro Analytical Balance as per the specifications mentioned in Annexure and as per the terms and conditions detailed below:-

1. Technical bid containing the technical specification of the instruments/equipment shall be kept in a separate sealed envelope (1) and Price bid quoting the price of the instrument shall be kept separately in another sealed envelope (2) mentioning “Technical Bid” and “Price bid” on top left side of the envelope respectively.
2. Both Technical and Price bids are to be kept separately in sealed envelopes and both shall be kept in another sealed envelope(3) mentioning that “Quotation for Micro Analytical Balance” on top left side of the envelope.
3. Separate quotation shall be submitted for each and every instrument/equipment individually depending on capacity to supply the instrument as per the terms and conditions.
4. Combined submission of the quotations for all the instrument shall be summarily rejected.

Contd....2/-

5. The duly filled in quotation shall reach this office on or before 18th March 2014 at 16:00 hours by post. The quotation may also be dropped in the Tender Box kept at this office.
6. The quotation received without separate sealed envelopes for technical and Price bid as explained in conditions 1 and 2 shall not be entertained.
7. The instrument shall be delivered, installed and adequate demo & training to the laboratory staff shall be given at this office.
8. On site warranty for three years to be given.
9. The call of quotation does not bind the Central Board to place the order.
10. The quotations submitted in response to this notice may be rejected without assigning any reason.
11. The instrument shall be delivered along with the invoice in original at this office within 30 days from the date of issue of Supply Order
12. The validity of quotation shall normally be 90 days from the date of opening.
13. The payment shall be released after the delivery, demo and satisfactory performance of the instrument.
14. Applicable taxes, cess and any other charges if any shall be quoted separately.

Yours faithfully

Sd...../-
(S.Suresh)
ZONAL OFFICER

Encl.: Annexure

Copy to

1. I/c. IT Divn., CPCB, HO. With a request to place in the official website
2. Notice Board, CPCB, SZO., Bangalore
3. Notice Board, KSPCB, Bangalore

Yours faithfully

(S.Suresh)
ZONAL OFFICER

ANNEXURE**Micro Analytical Balance (5.1g)**

S. No.	Specifications	
1.	Capacity	5.1 g
2.	Readability	1 µg
3.	Repeatability	± 1 micrograms
4.	Linearity, micrograms	± 4 micrograms
5.	Linearity	0 -500 µg ± 2 micrograms
6.	Operating Temperature	5° C to 40 °C,
7.	Operating Humidity	85% RH or less
8.	Pan Size (mm / in)	Ø 30 mm / 1.2 in Std pan Ø 50 mm / 1.9 in Std filter
9.	Automatic zero tracking	Std feature (can be turned off by menu code)
10.	Calibration	Self Calibration
11.	Power Source	230 ± 10 V AC with AC adopter with scale
12.	Standard Equipment	230 ± 10 V AC Adapter, level indicator, adjustable feet, RS-232, manual
13.	Warranty	Minimum 3 years onsite warranty
14.	Calibration certificate	From Accredited agency with one year validity.