

**Central Pollution Control Board**  
**Zonal Office, South**  
**7<sup>th</sup> D main, Thimmaiah Road**  
**Shiva Nagar, Bengaluru – 560010**  
**Phone: 080 23233739, FAX: 080-23234059**  
**E mail: cpcbszo@yahoo.com**

ZOB/Tech-42/Lab/2014-15/

13th November, 2014

To

**Sub: Inviting sealed quotation for purchase of dehumidifier**

Sir,

Sealed quotations are invited from reputed manufacturers /suppliers for supply of dehumidifier as per the specifications given in annexure-1. The terms and conditions of purchase are given below. Quotations should be submitted to this office **on or before 12<sup>th</sup> December, 2014**.

**Terms and Conditions:**

- a. The price bid shall be kept in a sealed envelope mentioning **“Quotation for supply of ‘Dehumidifier’ on top left side of the envelope.**
- b. The validity of quotation shall be for 90 days from the date of opening
- c. Rates should be quoted along with VAT no. and others, if any.
- d. Items which meet the specification only will be accepted.
- e. All charges including basic price, VAT, any other tax, and transportation cost etc. should be mentioned clearly.
- f. Duly filled in quotation shall reach this on or before **12<sup>th</sup> December, 2014 at 16.00 hours**. The quotations may be sent by registered post or may also be dropped in the tender box kept at this office.
- g. Payment will be made through A/C payee cheque in the name of organisation, after satisfactory supply of the item.
- h. Competent authority in the Board has full right to accept or reject any/all tenders /quotation
- i. With-out assigning any reason thereof.
- j. Typographical and clerical errors are subject to correction
- k. In the event of any dispute, the decision of the Chairman, CPCB, will be final binding upon both the parties.

Yours faithfully,

(S.Suresh)  
Zonal Officer

**Central Pollution Control Board**  
**Zonal Office, South**  
**7<sup>th</sup> D main, Thimmaiah Road**  
**Shiva Nagar, Bengaluru – 560010**  
**Phone: 080 23233739, FAX: 080-23234059**  
**E mail: cpcbszo@yahoo.com**

**Specifications for dehumidifier:**

Power	= < 180W
Extraction	10 – 12 lit /day
Coverage	25 – 30 m <sup>2</sup>
Tank capacity	1.5 – 1.8 lit
Running temperature	5 - 35°C
Compressor	Rotary
Coolant	To be free from CFC
Air flow	140-150 m <sup>3</sup> /hr

The dehumidifier should be compact and handy and should have features for silent operation, auto humidity sensor, LCD display of temperature and humidity, facility for humidity settings, continuous drainage system, pre-filter screen.

The equipment shall be supplied with all necessary accessories including charcoal filter, PVC hose, manual, necessary electrical fittings etc. The equipment shall be complete in all aspects.